



GOVERNMENT COLLEGE UNIVERSITY FAISALABAD

TENDER DOCUMENT No. NRPU/7345/2019

SUPPLY OF EQUIPMENT FOR DEPARTMENT OF BIOCHEMISTRY, GCUF.

Purchase Price : Rs. 500/-

Sealed Tenders are invited from the well reputed firms / suppliers, who are registered with Tax Departments for Purchase of Equipment for Department of Biochemistry, GCUF.

GENERAL TERMS & CONDITIONS

1. Tender Opening Date & Procedure:

The procurement shall be completed in accordance with Punjab Procurement Rules 2014, on Single Stage - Two Envelope Bidding Procedure. Bids in complete conformity with Tender Documents will be dropped in Tender Box placed at the office of undersigned, not later than **18-10-2019 on 10:00 A.M** which shall be opened on same day at **10:30 A.M** in Treasurer's Office.

Note: Tender Number must be mentioned on envelope.

2. Tender Fee, Bid Security and Performance Security:

The Bid must be accompanied by Tender Fee of Rs:500/- in shape of Call Deposit Receipt (CDR) in original and Bid Security of 2% of Estimated Price of Rs.998,700/- (refundable) in shape of Call Deposit Receipt (CDR) in original. CDRs must be in favor of Treasurer, Government College University, Faisalabad. The rates / bids should be inclusive of all applicable Govt. Taxes.

3. Validity of Offers.

- i. Offers shall be valid for 120 days from the date of submission of bids.
- ii. Withdrawal / modification of the original offer within the validity period shall entitle the University to forfeit Bid Security.
- iii. Sales Tax Number and NTN Number must be mentioned on the letter head and bidding documents.

4. Failures and Terminations:

No offer of a supplier / firm will be considered if:-

- a) Bid received without Bid Security / Call Deposit or less CDR than the required.
- b) Bids received not in accordance with specifications of Tender Documents.
- c) Bid received later than the date and time fixed for tender.
- d) Tender is incomplete in any respect or is unsigned.
- e) Offer is ambiguous and the offer is conditional.
- f) Offer from a firm which is black listed at any level.
- g) Any erasing / cutting / overwriting etc.
- h) The supplier fails to deliver the consignment within specified delivery period strictly in accordance with the terms and conditions as laid down in the Purchase Order.
- i) Situation warranted, then University is authorized to forfeit the bid Security and the firm may also be black listed.

5. Other special conditions

- I. The decision of the Vice Chancellor of Government College University, Faisalabad, would be final & binding on both the parties and not challengeable in any court of law.
- II. 10% performance security under the rules shall be deducted from the final bill of the successful bidder as per PPRA Rules 2014. (If applicable)
- III. All Government Taxes will be deducted according to applicable rules.
- IV. The University may reject all bids/proposals at any time prior to the acceptance of a bid or proposal as per PPRA Punjab rule 35, however upon bidder request the ground of rejection will be communicated to the concerned but no justification will be given as per PPRA rule 35 (2).

Prof. Dr. Shazia Anwer Bukhari
Department of Biochemistry
GC University Faisalabad
Cell: 0300 6637159

ISSUED TO:

M/S-----

TENDER DOCUMENT No.NRPU/7345**SUPPLY OF EQUIPMENT FOR DEPARTMENT OF BIOCHEMISTRY, GCUF.**

Sr.No.	Items	Specification or Equivalent	QTY	Unit Price Including Taxes	Total Cost
1.	<u>CO2 INCUBATOR</u>	<u>Specifications:</u> Compact CO₂ Incubator should be operations with animal cell cultures and tissues, DNA/RNA reaction analysis, hybridization reactions, synthesis of target proteins, other molecules and synthesis of antibodies, other proteins of immune system. <u>UV Lamp</u> Should be equipped with a UV air recirculation system – 1 UV lamp and a fan. Should be Six-sided heating provides uniform distribution of the temperature inside the u, Infrared CO₂ sensor, non-sensitive to temperature and humidity changes, Bluetooth data transfer to PC, Error tracing and alarm system and separately heated lockable port for chamber access for cables. Working chamber material Stainless Steel (1 mm), Temperature setting range +25°C ... +60°C, Temperature stability ±0.1°C, Temperature uniformity at +37°C, ±0.3°C, Working volume 46 litres, Number of shelves 3, Presence of the inner glass door +, Relative humidity >90% @ 37°C, Humidity delivery Water Pan, CO₂ control range* 0 – 20%, *At set temperature from ambient to 50 °C, +, CO₂ sensor Infrared CO₂ sensor, Temperature and CO₂ level input Digital, UV lamp 1 × 6 W, TUV G6T5, Data output Wireless, Access port 1 (ø 26 mm), Internal chamber dimensions 350 × 310 × 385 mm, Overall dimensions (W×D×H) 400 × 410 × 580 mm, Power consumption 600 W and Nominal	01		

		operating voltage 230V, 50/60 Hz; 115 V, 50/60 Hz. EXTRA Incubator should be Chamber capacity 32 liters, Natural air convection, Microprocessor temperature controller, Temperature sensor fail alarm, Power failure control system, Real time clock. Should be Solid door, RS 232 interface and USB port, Stainless steel wire shelves, Manufacturer's test certificate at +37°C Open door alarm, Air-flap diameter 37 mm. Temperature Range 5° above ambient temperature ...+100°C, Nominal Power [W] 300, Shelves Fitted 1 and Voltage 230 V. EXTRA Vertical Gel Electrophoresis Unit Number of Gels 1-4, Plate Dimensions (W x L): 10 x 10 cm, Gel Dimensions (W x L): 8 x 8.5 cm, Max Samples Capacity: 80 samples, 20 samples per gel, Buffer Volume for 2 Gels: Min 250 ml; Max 1200 ml and Buffer Volume for 4 Gels: Min 250 ml; Max 1200 ml. <u>Make: Europe / USA</u>			
--	--	---	--	--	--

Note: (1) Rate should also be quoted on Firm's letter head pad.

(2) China/Korea made instruments/equipment will not be accepted

Details of Call Deposit Receipt (CDR)

Signature _____

No _____ **Amount** _____

Name of Bidder _____

Bank _____ **Branch** _____ **City** _____

CNIC No _____

Name of the Firm & Address _____ **Ph No** _____

N.T.No. _____ **Sales Tax No.** _____

Issued By:

**Prof. Dr. Shazia Anwer Bukhari
Department of Biochemistry GC University,
Faisalabad**